



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

| | |
|-------------------------|-----------------------|
| RPC-N 75 Ω according to | IEC 61169-16 |
| F 75 Ω according to | IEC 61169-24, EIA-550 |

Documents

| | |
|------------------|------------------------------|
| Application note | AN001 "Calibration Services" |
|------------------|------------------------------|

Material and plating

Connector parts

| | | |
|----------------|-------------------------|---|
| Center contact | Material CuBe | Plating Gold, min. 1.27 µm, over nickel |
| Outer contact | Stainless steel | Passivated |
| Coupling nut | Stainless steel | Passivated |
| Dielectric | PS | |

Electrical data

| | |
|-------------|---|
| Frequency | DC to 6 GHz |
| Return loss | ≥ 32 dB, DC to 3 GHz ≥ 28 dB, 3 GHz to 6 GHz |

Mechanical data

| | | |
|----------------------|--------------------|--------------------|
| | RPC-N 75 Ω | F 75 Ω |
| Mating cycles | ≥ 500 | ≥ 1000 |
| Maximum torque | 1.70 Nm | 6.78 Nm |
| Recommended torque | 1.10 Nm | 2.00 Nm |
| Gauge | 5.28 mm to 5.36 mm | 0.00 mm to 0.10 mm |
| Nominal pin diameter | | 0.81 mm |

General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

| | |
|-------------------------------------|----------------|
| Offset Z_0 / Impedance / Z_0 | 75 Ω |
| Offset Delay | 143.099 ps |
| Length (electrical) / Offset Length | 42.90 mm |
| Offset Loss | 3.20 GΩ/s |
| Loss | 0.0265 dB/√GHz |
| Line Loss @ 1GHz | 0.0006 dB/mm |

Environmental data

| | |
|---|-------------------|
| Operating temperature range ¹ | +20 °C to +26 °C |
| Rated temperature range of use ² | 0 °C to +50 °C |
| Storage temperature range | - 40 °C to +85 °C |

RoHS compliant

¹ Temperature range over which these specification are valid.

² This range is underneath and above the operating temperature range, within the calibration adaptor is fully functional and could be used without damage.

Technical Data Sheet

Rosenberger

Calibration Adaptor
RPC-N 75 Ω Plug – F 75 Ω Plug

P5S174-S21S3

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Packing

Standard 1 pce in box
Weight 48.0 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

| Draft | Date | Approved | Date | Rev. | Engineering change number | Name | Date |
|--|----------|---------------|----------|---|---------------------------|------------------|---------------|
| Marcel Panicke | 08.08.18 | Markus Müller | 09.05.19 | b00 | 19-0897 | Marion Striegler | 09.05.19 |
| Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de | | | | Tel. : +49 8684 18-0 Email : info@rosenberger.de | | | Page 3 / 3 |