
THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

REVISION						
REV.	DESCRIPTION	DATE	APPROVED			





CONTROLLED

NOTES:

1.	CORDAGE: SJT, 18AWG/3C, UNSHIELDED, NA COLOR CODE, TEMP. RATING 60°C, RATING: 125V, 10A, JACKET COLOR: GRAY
2.	PLUG: NEMA 5-15P HOSPITAL GRADE (CLEAR)
3.	CONNECTOR: IEC 60320 C-13 RIGHT ANGLED (GRAY)
4.	APPROVALS: UL, CSA
5	ROHS COMPLIANT

- ROHS COMPLIANT 5. 6.
- CAUTION LABEL

ſ	Dim range	Tolerance +/- (mm)	Qualtek Electronics Corp. PPC DIVISION			
	(mm)		Part Number:	240003-06	RoH5 Compliant	
[0-1	0.15	Description: POWER CORD HOSPITAL GRADE			
	1-6	0.20	POWER CORD HOSPITAL GRADE			
	6-18	0.30				
ľ	18-50	0.50		UNIT: in		
	50-120	0.70	Drawn / Date	Checked / Date	Approved / Date	
	120-250	0.90	R.Nuti	01-10-08	JJH / 01-10-08	
	250-500	1.20	01-10-08		00-01-10 /1166	