


The logo consists of the letters 'B+B' in a white, serif font, centered on a red rectangular background.

SENSORS



1. Plug in the Mineral Insulated Thermocouple NiCr-Ni, Type K with Miniature Plug into the destined socket at the measurement device.
2. Read out the measured value of the measurement device.
3. Measuring range of Mineral Insulated Thermocouple $-200^{\circ}\text{C} \dots +1100^{\circ}\text{C}$
($-328^{\circ}\text{F} \dots +2012^{\circ}\text{F}$)

Operating Instruction

Mineral Insulated Thermocouple
with miniature plug

B+B Thermo-Technik GmbH | Heinrich-Hertz-Straße 4 | D-78166 Donaueschingen
Fon +49 771 83160 | Fax +49 771 831650 | info@bb-sensors.com | bb-sensors.com