


SHEET 2.1

CONSTANT VOLTAGE DRIVE CONDITIONS



FREQUENCY IN HERTZ

ACOUSTICAL

SENSITIVITY

DEVICE WILL PRODUCE THE SPL LISTED BELOW UNDER TEST CONDITIONS DESCRIBED IN TABLE 3. NOMINAL SENSITIVITY AT IKHZ IS dB RELATIVE TO 20μPα. ALL OTHER VALUES IN dB RELATIVE TO THE SENSITIVITY AT IKHz.

FREQUENCY (Hz)	MINIMUM	NOMINAL	MAXIMUM
200	+1.5	+3.5	+5.5
500	0.0	+2.0	+4.0
1000	-2.0	102.5	+2.0
2000 - 3000	-1.0	+2.0	+5.0
4000	-12.0		

TABLE I

TOTAL HARMONIC DISTORTION

DEVICE WILL NOT EXCEED TOTAL HARMONIC DISTORTION LEVELS LISTED BELOW.

FREQUENCY (Hz)	AC DRIVE (V rms)	DC BIAS (mA)	LIMIT (%)
500	0.470	0	10
830	0.167	0	5
1250	0.167	0	5

TABLE 2

TEST CONDITIONS

NOMINAL SOURCE VOLTAGE	0.167 V rms, OmA DC BIAS
SOURCE IMPEDANCE	< I Ohm
TUBING	
COUPLER CAVITY	2 CM ³ , SIMULATED ANSI S3.7 TYPE HA-3 (IEC 126)

TABLE 3

ELECTRICAL

DC RESISTANCE	48 OHMS ± 10%
IMPEDANCE @ 500 Hz	65 OHMS \pm 15%
IMPEDANCE @ I kHz	100 OHMS ± 15%

TABLE 4

ISOLATION: CASE WILL BE ELECTRICALLY ISOLATED FROM THE COIL CIRCUIT.

TEMPERATURE: OPERATING RANGE FROM 0°C TO 63°C (SENSITVITY WILL NOT VARY BY MORE THAN ±3 dB WITHIN RANGE)

SENSITIVITY AT 0°C IS 2dB LOWER THAN THE SENSITIVITY AT ROOM TEMPERATURE. DELTA PEAK IS IdB HIGHER AT BODY TEMPERATURE (37°C)

STORAGE RANGE FROM -40°C TO 63°C.

Revision	C.O. #	implementation vare	KELEASE LEVEL		KEN1210N
С	C10107990	11-20-08			\sim
В	C10104212	5-1-06	Active		(
А	C10103609	2 - 2 - 06)
WHEN TEST LIMITS ARE USED TO ESTABLISH INCOMING INSPECTION ACCEPTANCE/REJECTION			DR. BY	DATE	
		OF TEST EQUIPMENT WITH KNOWLES IS ALSO REQUIRED FOR ENT AND TEST METHOD VARIATION		CRG	2 - 2 - 06
				CK. BY	DATE
TITLE:	RF	CFIVER	FED-26792-102	GJP	2-8-06
	- ' ' '		. 25 25, 62 162	APP. BY	DATE

SHT 2.1

GJP

2-8-06

PERFORMANCE SPECIFICATION

PORT LOCATION: 12C

KNOWLES ELECTRONICS
ITASCA, ILLINOIS U.S.A.