GlobTek, Inc. www.globtek.com	Drawing Title: DANISH/DENMARK HOSPITAL GRADE PLUG, MEDICAL GRADE PLUG, MALE PLUG AFSNIT SECTION 107-2-D1 (DK 2-8A) to IEC60320 C13 2 CONDUCTORS + GR 1.0 mm² / 3 Conductors, 7.2 ± 0.2 mmmm OD, H05VV-F, 70°C Deg.C, 250V, BLU/BRN/GRN-YW, Approvals: IEC 60227, 2500mm, GlobTek Black	ROUND, GT Cable #427-T:
	Model No. Power Cord	
Northvale, NJ 07647		
Tel. (201) 784-1000	Part No./Ordering No. G801M4272M5701(R)	Rev A
Fax (201) 784-0111		

In Addition to GlobTek Inc.'s renewed ISO9001:2008 - Quality Management System Certification, GlobTek Inc. is now certified to:

ISO13485:2003 - Medical Devices Quality Management System Certification

ISO14001:2004 - Environmental Management System Certification

ISO Certificates are available online at https://www.globtek.com/iso-certificates/

Prop 65 Statement:https://en.globtek.com/prop-65-statement.pdf

The acceptance of this specification and use of the product described in this document indicates the acceptance and binding of the customer to Globtek TERMS AND CONDITIONS (attached to the end of this document), which supersede all other agreements, terms, and conditions (https://en.globtek.com/terms-and-conditions/).

The purchaser of this product (Customer) agrees that it is their sole responsibility to determine suitability/fitness of the product described in this specification for their application and to take precautions for protection against any hazards attendant to the handling and use of the product.

Footnote:

GlobTek Inc. will not be liable for the safety and performance of these power supplies if unauthorized access and repair occurs. End user should consult applicable UL, CSA or EN standards for proper installation instruction.

Limitation of Use:

GlobTek products are intended for use with commercial, medical, and/or industrial equipment only. GlobTek product are not authorized for use as mission critical components in life support hazardous environment, nuclear or aircraft applications without prior written approval from the CEO of GlobTek Inc.

Contents of this document are subject to change without prior notice.

Customer Approval of Specification:

Please approve, sign and send back to GlobTek so we can complete order processing.

A delay in receipt of this form will delay delivery schedule.

Company Name:	Customer P/N:
Quote Number:	Date:
Authorized Representative Name:	Authorized Representative Signature:

GlobTek, Inc. www.globtek.com	Drawing Title: DANISH/DENMARK HOSPITAL GRADE PLUG, MEDICAL GRADE PLUG, MALE PLUG AFSNIT SECTION 107-2-D1 (DK 2-8A) to IEC60320 C13 2 CONDUCTORS + GR 1.0 mm² / 3 Conductors, 7.2 ± 0.2 mmmm OD, H05VV-F, 70°C Deg.C, 250V, BLU/BRN/GRN-YW, Approvals: IEC 60227, 2500mm, GlobTek Black	ROUND, GT Cable #427-T:
•	Model No. Power Cord	
Northvale, NJ 07647		
Tel. (201) 784-1000	Part No./Ordering No. G801M4272M5701(R)	Rev A
Fax (201) 784-0111		

REVISION HISTORY

	REV	DESCRIPTION	SUB	DATE	APPROVED
Ш	А	Initial release	DR	10/19/2021	DAVID


Cordset Information:

CordsetColor:GlobTek Black

Cord Length (mm):2500

NOTES:

GlobTek, Inc. www.globtek.com Drawing Title: DANISH/DENMARK HOSPITAL GRADE PLUG, MEDICAL GRADE PLUG, MALE PLUG AFSNIT SECTION 107-2-D1 (DK 2-8A) to IEC60320 C13 2 CONDUCTORS + GROUND, GT Cable # 1.0 mm² / 3 Conductors, 7.2 ± 0.2 mmmm OD, H05VV-F, 70°C Deg.C, 250V, BLU/BRN/GRN-YW, Approvals: IEC 60227, 2500mm, GlobTek Black			
	Model No. Power Cord		
Northvale, NJ 07647			
Tel. (201) 784-1000	Part No./Ordering No. G801M4272M5701(R)	Rev A	
Fax (201) 784-0111			



PROPRIETARY INFORMATION

GlobTek, Inc. www.globtek.com Drawing Title: DANISH/DENMARK HOSPITAL GRADE PLUG, MEDICAL GRADE PLUG, MALE PLUG AFSNIT SECTION 107-2-D1 (DK 2-8A) to IEC60320 C13 2 CONDUCTORS + GROUND, GT Cable 1.0 mm² / 3 Conductors, 7.2 ± 0.2 mmmm OD, H05VV-F, 70°C Deg.C, 250V, BLU/BRN/GRN-YW, Approvals: IEC 60227, 2500mm, GlobTek Black			
	Model No. Power Cord		
Northvale, NJ 07647			
Tel. (201) 784-1000	Part No./Ordering No. G801M4272M5701(R)	Rev A	
Fax (201) 784-0111			

CUSTOMIZED LENGTHS, COLORS, CABLE TYPES, TERMINATIONS/PLUGS/RECEPTACLES AVAILABLE UPON REQUEST. ADDITIONAL PLUG OPTIONS ARE AVAILABLE AT https://www.globtek.com/connectors/ STANDARD PLUGS AND CONNECTORS CAN BE SEEN ONLINE AT https://www.globtek.com/connectors/ ACTUAL CONNECTORS, OVERMOLDS, FERRITES MAY VARY SLIGHTLY FROM THE PICTURE BELOW DIMENSIONS ARE IN MM UNLESS SPECIFIED OTHERWISE

End A	Cable	End B
Code:G801M	Code:427-T	Code:701-ENEC
Description: DANISH/DENMARK HOSPITAL GRADE PLUG,	Description:GT Cable #427-T: 1.0 mm² / 3 Conductors, 7.2 ±	Description:IEC60320 C13 2 CONDUCTORS + GROUND
MEDICAL GRADE PLUG, MALE PLUG AFSNIT SECTION	0.2 mmmm OD, H05VV-F, 70°C Deg.C, 250V,	Rating (V):250
107-2-D1 (DK 2-8A)	BLU/BRN/GRN-YW, Approvals: IEC 60227	Rating (A):10.0
Rating (V):250	Color:GlobTek Black	Gender:Female
Rating (A):10.0	Wire Gauge(AWG):1.0 mm²	Overmolded:Yes
Gender:Male	Cond:3	
Overmolded:Yes	Type:H05VV-F	
	Temperature:70°C° C	
	Amps:10	
	Volts:250	
	OD:7.2 ± 0.2 mm	
	ConductorsColors:BLU/BRN/GRN-YW	

Overall cable rating:10 / 250 v

lug Approvals		Cable Approvals		Connector Approvals	
(\mathbf{D})	Demko	\$ \$ 0 0 \$ 0 ∆	European approvals	10 <u>P</u> VE	ENEC-VDE
AGENCY ONLINE DOCUMENTS	http://er	n.globtek.com/cordsets/			
RoHS/RoHS3 Declaration	https://v	vww.globtek.com/pdf/rohs_	cert/01t3a000005JqinAAC		
REACH Declaration	https://v	vww.globtek.com/pdf/iso_co	ertificates/REACH.pdf		
Conflict Minerals Declaration	https://v	vww.globtek.com/pdf/confli	ct-minerals.pdf		
Prop 65 Statement	https://e	n.globtek.com/prop-65-sta	tement.pdf		
CE.UKCA.Cord	https://g	lobtek.secure.force.com/sp	pec/ec_decl_product_model_family	y_incl_cord?id=01t3a000005Jqin	


PROPRIETARY INFORMATION

GlobTek, Inc. www.globtek.com	ROUND, GT Cable #427-T:	
186 Veterans Drive,	Model No. Power Cord	
Northvale, NJ 07647		
Tel. (201) 784-1000	Part No./Ordering No. G801M4272M5701(R)	Rev A
Fax (201) 784-0111		

Hanking:200

Description:200 +/- 20mm

Image:



Notes/Modifications/Customizations (notes in this section supersede the values shown in the standard spec):