# NEVO+600ML LOW NOISE MEDICAL DATA SHEET

AC/DC Modular Configurable PSU





### 600 Watts in the palm of your hand

The NEVO+600ML modular configurable medical power supply is the smallest in its class and the ultimate solution for demanding medical applications where size, power density and weight matter. Its tiny footprint of 5" x 3" x 1.61" weighs only 600 grams and delivers an incredible 450 Watts with a minimum of audible noise. The input module can accommodate up to four isolated output modules which can be configured into a high power 5"x 3" single output power supply or a multiple output power supply with up to 8 isolated outputs. Standard features include intelligent fan control providing optimised airflow for various load and temperature conditions, wide output voltage adjust, parallel and series connection of modules and an isolated 5V 1A bias supply. The low noise fan option allows you to use this innovative power supply in even the quietest of environments. The series is approved to latest medical standards and features market leading specifications and design in application support.

### MAIN FEATURES

• 450 Watts in 5" x 3" x 1.61"	<ul> <li>Up to 8 isolated outputs</li> </ul>	<ul> <li>Accurate current sharing</li> </ul>
<ul> <li>Low noise operation (~18dBA reduction from S version)</li> </ul>	User and field configurable	<ul> <li>Standard 5V 1A bias supply</li> </ul>
Intelligent fan control	<ul> <li>Parallel and series connection of modules</li> </ul>	<ul> <li>IEC/UL60601-1 Ed. 3 &amp; -1-2 Ed. 4 (EMC)</li> </ul>
<ul> <li>Efficiency up to 89%</li> </ul>	<ul> <li>Wide output voltage adjust range</li> </ul>	• 3 Year warranty
	<ul> <li>Remote current / voltage programming</li> </ul>	<ul> <li>Parallel units with droop current sharing</li> </ul>
APPLICATIONS		
Medical & diagnostic equipment	<ul> <li>Telecommunications</li> </ul>	Lasers
<ul> <li>Test &amp; Measurement equipment</li> </ul>	<ul> <li>Laboratory &amp; Analysis equipment</li> </ul>	LED lighting
Robotics	• Display	<ul> <li>Retrofit of legacy PSUs</li> </ul>
• Oil & Gas	Avionics	
CUSTOMER BENEFITS		
Fast time to market	<ul> <li>Proven technology</li> </ul>	<ul> <li>Technology consolidation</li> </ul>
<ul> <li>24 hrs samples from distribution</li> </ul>	<ul> <li>Eliminates custom design costs</li> </ul>	<ul> <li>Supplier consolidation</li> </ul>
<ul> <li>Safety &amp; EMC certified</li> </ul>	<ul> <li>Field replaceable</li> </ul>	
<ul> <li>World class engineering support</li> </ul>		

Page 1 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com

## **SPECIFICATIONS**

INPUT MODULE SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
AC Input Voltage	Nominal range is 100V <sub>RMS</sub> to 240V <sub>RMS</sub>	85		264	V <sub>RMS</sub>	
AC Input Frequency	Contact factory for 400Hz operation.	47	50/60	63	Hz	
DC Input Voltage	Not covered by safety approvals. Contact Vox Power.	120		300	V <sub>DC</sub>	
Output Power Rating	De-rate linearly from 450Watts at 120V <sub>RMS</sub> to 338Watts at 85V <sub>RMS</sub>			450	Watts	
Input Current	450Watts output at 120 V <sub>RMS</sub> input			5	Amps	
Input Current Limit	Maintains power factor		8		Amps	
Inrush Current	265V <sub>RMS</sub> , 25°C (cold start)			20	Amps	
Fusing	Live line fused (5x20 Fast acting)			8	Amps	
Efficiency	See graphs		86	89	%	
No load Power consumption	All outputs fitted and disabled/enabled		21/28		Watts	
Power Factor	Typical value for 300 Watts output at 240Vrms input		0.96	0.99		
Holdup	450Watts output at 120V <sub>RMS</sub> input	17	20	21	mS	
UVP	Turn on under voltage protection	78		84	V <sub>RMS</sub>	
Over temperature	Internally monitored.	115		125	°C	
Reliability (1)	Input module			1.207	FPMH	
	Fan			2.7	FPMH	
Warranty	Standard terms and conditions apply			3	Years	
Size	133.7 (L) x 77.7 (W) x 41.0 (H). See diagram for tolerance details				mm	
Weight	360 + 60 per output module				Grams	
Note 1.	30°C base & ambient, 100% load, SR332 Issue 2 Method I, Case 3, Ground, Fixed, Control	lled				

GLOBAL SIGNALS SPECIFICATIONS					
Parameter	Details	Min	Typical	Max	Units
Bias Voltage	Two isolated Bias Outputs available	4.8	5	5.2	Volts
Bias Current	Hiccup type current limit	0		1	Amps
AC_OK Voltage	Low output level High output level	0 3.5	0.2 4.5	1 5.2	Volts
AC_OK Current		-10		20	mA
Power Good Voltage	Low output level. internal 10kΩ pull down. High output level. PNP open collector.	0 8	0 10	0 15	Volts
Power Good Current	Open collector output. Current source only. All Slots.			20	mA
Global Inhibit Voltage	Low input level High input level	0 3		1 15	Volts
Global Inhibit Current	5k input impedance.	0.6		3	mA
Inhibit Voltage	Low input level. All slots. High input level. All slots.	0 2.5		1 15	Volts
Inhibit Current	10k input impedance. All slots.	0.25		1.5	mA

	OUTPUT MODULE SPECIFICATION SUMMARY											
MODEL	Output Voltage		age	Output	Rated	Peak	Load	Line	Cross	Ripple &	FPMH <sup>(1)</sup>	Feature
MODEL	Min.	Nom.	Max.	Current	Power	Power	Reg.	Reg.	Reg.	Noise		Set (2)
OP1	1.5V	5V	7.5V	25A	125W	187.5W	±50mV	±5mV	±10mV	50mV <sub>PP</sub>	0.5	ABCDEFG
OP2	4.5V	12V	15V	15A	150W	225W	±100mV	±12mV	±24mV	120mV <sub>PP</sub>	0.5	ABCDEFG
OP3	9V	24V	30V	7.5A	150W	225W	±150mV	±24mV	±48mV	$240 mV_{PP}$	0.5	ABCDEFG
OP4	18V	48V	58V	3.75A	150W	217.5W	±300mV	±48mV	±96mV	480mV <sub>PP</sub>	0.5	ABCDEFG
OP5	3.3V	12V	15V	5A	2x 75W	2x 75W	±50mV	±12mV	±24mV	240mV <sub>PP</sub>	0.75	AFG
OP8	23.2V	24V	24.7V	3.125A	2x 75W	2x 75W	±100mV	±24mV	±48mV	480mV <sub>PP</sub>	0.75	AFG
OPA2 <sup>(3)</sup>	4.5V	12V	15V	25A	300W	375W	±100mV	±12mV	±24mV	120mV <sub>PP</sub>	0.5	ABCDEFGH
OPA3(3)	9V	24V	30V	15A	300W	450W	±150mV	±24mV	±48mV	240mV <sub>PP</sub>	0.5	ABCDEFGH
Note 1. Note 2.	Note 1. Output module, 30°C base, 100% load, SR332 issue 2 Method I, Case 3, Ground, Fixed, Controlled											

G = Over temperature protection, H = Dual Slot module

Note 3. Can only be used with NEVO+600 chassis with date codes from 2048 onwards. eg. 2048C080000 can use A2 or A3 module, 2047C089999 cannot use A2 or A3 module.

	SAFETY SPECIFICATIONS					
Parameter	Details	Max	Units			
	Input to Output (2 MOPP). Do not perform test on assembled unit <sup>(1)</sup>	4000	V <sub>AC</sub>			
Isolation Voltages	Input to Chassis (1 MOPP)	1500	V <sub>AC</sub>			
	Global signals (J2) to Output/Chassis	250	V <sub>DC</sub>			
	Output to Output/Chassis (Standard modules)	250	V <sub>DC</sub>			
Earth Leakage Current	Normal condition, 264Vac, 63Hz, 25°C	300	uA			
Touch Leakage Current	Standard modules NC/SFC	20/200	uA			
Patient Leakage Current	Standard modules 264Vac, 63Hz, 25°C NC/SFC <sup>(2)</sup>		uA			
Note 1. Testing an assembled unit to 4000V <sub>AC</sub> may cause damage. Please refer to application note (APN-002) on Vox Power website or contact Vox Power representative.						
Note 2. Not Applicable						


INSTALLATION SPECIFICATIONS						
Parameter Details Parameter Details						
Equipment class		Flammability Rating	94V-2			
Overvoltage category	II	Ingress protection rating	IP10			
Material Group	IIIb (indoor use only)	ROHS compliance	2011/65/EU & 2015/863/EU			
Pollution degree	2	Intended usage environment	Home Healthcare			


Page 2 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com

			ENVIRONMENTAL SPECI	FICATIC	DNS				
<b>D</b> .				٩	lon-Op	erational	Opera	itional	11
Parameter	Details				Min Max		Min	Max	Units
Air Temperature	Operational limits subject to	appropri	iate de-ratings		-40	+85	-20	70	°C
Humidity	Relative, non-condensing				5	95	5	95	%
Altitude					-200	5000	-200	3000	m
Air Pressure					52	106	69	106	kPa
Noise Level	Variable. Measured 1m from				-	-	18	42	dBA
Shock	3000 bumps at 10G (16ms) h								
Vibration	1.5G 10 to 200Hz sine wave,		5min in 3 axes random vibration						
		ELEC	CTROMAGNETIC COMPLIAN	NCE – E	MISSIC	DNS			
Phenomenon			Basic EMC Standard		Tes	t Details			
Radiated emission	s, electric field		EN55011/22, FCC		Clas	s B compliant			
Conducted emission	ons		EN55011/22, FCC part 15, CISPR 22/1	1	Clas	s B compliant			
Harmonic Distortic			IEC61000-3-2	Compliant					
Flicker & Fluctuation IEC61000-3-3 Compliant									
		ELE	CTROMAGNETIC COMPLIA	NCE – II	MMUN	ITY			
Phenomenon			Basic EMC Standard		Tes	t Details			
Electrostatic discha	arge		IEC61000-4-2	Test level 4: 15kV air, 8kV contact					
Radiated RF EM fie	lds		IEC61000-4-3	Test Level 3: (10V/m, 80MHz-2.7GHz) sine wave AM 80% 1kHz				kHz	
Proximity fields fro equipment	m RF wireless communications		IEC61000-4-3	Test levels as per IEC60601-1-2:2014 Table 9					
Electrical Fast Tran	sients/bursts		IEC61000-4-4	Test Level 3: (2kV Power, 1kV I/O) 5kHz(ed3) & 100kHz(ed4)					4)
Surges			IEC61000-4-5	Test Level 3: 1kV L-N, 2kV L-E					
Conducted disturb	ances induced by RF fields		IEC61000-4-6	Test Level 3: 10V, 0.15 to 80Mhz sine wave AM 80% 1kHz					
Power Frequency I	Magnetic Fields		IEC61000-4-8	Test lev	vel 4: 30A	/m 50Hz			
Voltage Dips			IEC61000-4-11& SEMI-F47-0706 <sup>(2)</sup>	0% 10ms, 0% 20ms, 80% 1s, 80% 10s, 90% continuous (Criterion A) 70% 0.5s, 40% 0.2s (Criterion A at 240V and Criterion B at 100V)					
Voltage interruption	ons		IEC61000-4-11	0% 250	/300 cyc	e as per IEC60	501-1-2:2014	(Criterion B)	
Notes: 1.	Criterion C = Temporary loss of	adation of of function	nance or loss of function. f performance or loss of function is allo n is allowed but requires operator inte ). Line deratings applied where approp	rvention to			f-recoverable		
			AGENCY APPROV	/ALS					
Standard						File			
IEC 60601-1:2005 + + A1:2012	+ CORR1 2006 + CORR2: 2007	Medica perform	l electrical equipment Part 1: General r	equireme	nts for ba	sic safety and	essential	UL: E31	6486

+ A1:2012	performance	
EN60601-1:2006 + A11:2011 + A1:2013 +	Medical electrical equipment Part 1: General requirements for basic safety and essential	
A12:2014	performance	
CAN/CSA-C22.2 No. 60601-1 (2008)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential	
CAN/CSA-C22.2 NO. 00001-1 (2000)	Performance	
ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential	
ANSI/AANII ES00001-1 (2005 + C1.09 + A2.10)	Performance	
CE MARK	LVD 2014/35/EU, EMC 2014/30/EU	

CB certificate and report available on request






Page 3 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com

#### DOC-DTS-016-07, NEVO+600ML Medical Low Noise Datasheet

Temperature


### MECHANICAL DIMENSIONS AND MOUNTING SCREWS





#### CONNECTORS

0		<u>.</u>	
D	INI	$\sim$	

PINOUTS							
	J1						
Circuit	Details						
1	Live						
2	Earth						
3	Neutral						
	J2						
Circuit	Details						
1	Power good	Slot A					
2	Inhibit	SIOLA					
3	Power good	Slot B					
4	Inhibit	SIULD					
5	Power good	Slot C					
6	Inhibit	5101 C					
7	Power good	Slot D					
8	Inhibit	SIDED					
9	Global inhibit						
10	AC OK						
11	+5V 1A bias supply						
12	COM						
J5 <sup>(4)</sup>							
Circuit	Details						
1	-Sense						
2	+Sense						
3	Voltage control						









Positive output

Negative output

REF.	DETAILS	MANUFACTURER	HOUSING	TERMINAL					
J1	MAINS INPUT: 3 Pin, 5.08mm, with Friction Lock, 18-24 AWG	MOLEX	10013036	0008701031					
J2	GLOBAL SIGNALS: 12 Pin, 2mm, without Friction Lock, 24-30 AWG	MOLEX	511101251	0503948051					
J3/4 <sup>(1)</sup>	OUTPUT POWER TERMINAL: TAB SIZE 6.35mmx0.8mm	VARIOUS		VARIOUS					
J5	OUTPUT SIGNALS: 6 Pin, 1.25mm, with Friction lock, 28-32 AWG	MOLEX	0510210600	0500588000					
Notes									
1. Termin	1. Terminal and wire current rating must exceed maximum short circuit output current. Eg. Output 1 = 25A*1.25 = 31.25Amps								
2. Direct e	equivalents may be used for any connector parts								

3. All cables must be rated 105°C min, equivalent to UL1015

Current control / share / out

+5V 10mA local bias supply

4


5

COM

4

Pinout is for single output types only

#### PART NUMBERING SYSTEM



All specifications are believed to be correct at time of publishing. Vox Power Ltd reserves the right to make changes to any of its products and to change or improve any part of the specification, electrical or mechanical design or manufacturing process without notice. Vox Power Ltd does not assume any liability arising out of the use or application of any of its products and of any information to the maximum extent permitted by law. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any products of Vox Power Ltd. VOX POWER LTD DISCLAIMS ALL WARRANTIES AND REPRESENTATIONS OF ANY KIND WHETHER EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, IMPLIED WARRANTIES OF SUITABILITY, FITNESS FOR PURPOSE, MERCHANTABILITY AND NONINFRINGEMENT.

Please consult your local distributor or Vox Power directly to ensure that you have the latest revision before using the product and refer to the latest relevant user manual for further information relating to the use of the product. Vox Power Ltd products are not intended for use in connection with life support systems, human implantations, nuclear facilities or systems, aircraft, spacecraft, military or naval missile, ground support or control equipment used for the purpose of guidance navigation or direction of any aircraft, spacecraft or military or naval missile or any other application where product failure could lead to loss of life or catastrophic property damage. The user will hold Vox Power Ltd harmless from any loss, cost or damage resulting from its breach of these provisions.