



All dimensions are in mm; tolerances according to ISO 2768 m-H

**Interface**

According to

IEC 61169-4, EN 122190, DIN 47223

**Documents**

Application note

AN001 "Calibration Services"

**Material and plating**

**Connector parts**

Center conductor  
Outer conductor  
Dielectric

**Material**

CuBe  
Stainless steel  
PPE

**Plating**

Gold, min. 1.27 µm, over nickel  
Passivated

**Electrical data**

|                                       |                                                     |
|---------------------------------------|-----------------------------------------------------|
| Frequency range                       | DC to 8 GHz                                         |
| Return loss                           | ≤ 0.10 dB, DC to 4 GHz<br>≤ 0.15 dB, 4 GHz to 8 GHz |
| Error from nominal phase <sup>1</sup> | ≤ 1.0°, DC to 4 GHz<br>≤ 1.5°, 4 GHz to 8 GHz       |

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance.

**Mechanical data**

|                    |                    |
|--------------------|--------------------|
| Mating cycles      | ≥ 500              |
| Maximum torque     | 35 Nm              |
| Recommended torque | 2.26 Nm            |
| Gauge              | 1.78 mm to 1.82 mm |

**General standard definitions**

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

|                                     |                                |
|-------------------------------------|--------------------------------|
| Offset $Z_o$ / Impedance / $Z_o$    | 50 $\Omega$                    |
| Offset Delay                        | 87.394 ps                      |
| Length (electrical) / Offset Length | 26.20 mm                       |
| Offset Loss                         | 0.50 G $\Omega$ /s             |
| Loss                                | 0.0076 dB/ $\sqrt{\text{GHz}}$ |
| Short Inductance <sup>2</sup>       |                                |

<sup>2</sup> Short Inductances are determined individually for each Short circuit and are documented in a Calibration Certificate.

**Environmental data**

|                                             |                   |
|---------------------------------------------|-------------------|
| Operating temperature range <sup>3</sup>    | + 20 °C to +26 °C |
| Rated temperature range of use <sup>4</sup> | 0 °C to +50 °C    |
| Storage temperature range                   | - 40 °C to +85 °C |
| RoHS                                        | compliant         |

<sup>3</sup> Temperature range over which these specification are valid.

<sup>4</sup> This range is underneath and above the operating temperature range, within the calibration adaptor is fully functional and could be used without damage.

**Declaration of calibration options**

**Factory Calibration**

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to national / international standards. Model based standard definitions are individually optimized and reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

**Accredited Calibration**

Optional this calibration standard can be delivered with an Accredited Calibration (DAkkS) having the highest confidence in the traceability. The DAkkS Calibration Certificate issued reports individual calibration results in a complex format, traceable to national / international standards. Model based standard definitions are individually optimized and reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format as well as in a dense data set needed for data based standard definitions. The uncertainties are smaller than in a Factory Calibration.

*For further, more detailed information see application note AN001 on the Rosenberger homepage.*

**Calibration interval**

Recommendation 12 months

**Packing**

Standard 1 pce in box  
Weight 87 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

| Draft          | Date     | Approved    | Date     | Rev. | Engineering change number | Name             | Date     |
|----------------|----------|-------------|----------|------|---------------------------|------------------|----------|
| Marcel Panicke | 03.08.15 | Lars Ramtke | 05.08.16 | c00  | 16-1267                   | Marion Striegler | 05.08.16 |

|                                                                                                                                                        |  |  |  |                                                                                              |  |  |               |
|--------------------------------------------------------------------------------------------------------------------------------------------------------|--|--|--|----------------------------------------------------------------------------------------------|--|--|---------------|
| Rosenberger Hochfrequenztechnik GmbH & Co. KG<br>P.O.Box 1260 D-84526 Tittmoning Germany<br><a href="http://www.rosenberger.de">www.rosenberger.de</a> |  |  |  | Tel. : +49 8684 18-0<br>Email : <a href="mailto:info@rosenberger.de">info@rosenberger.de</a> |  |  | Page<br>3 / 3 |
|--------------------------------------------------------------------------------------------------------------------------------------------------------|--|--|--|----------------------------------------------------------------------------------------------|--|--|---------------|