NEVO+600M MEDICAL DATA SHEET

AC/DC Modular Configurable PSU





600 Watts in the palm of your hand

The NEVO+600M modular configurable medical power supply is the smallest in its class and the ultimate solution for demanding medical applications where size, power density and weight matter. Its tiny footprint of 5" x 3" x 1.61" weighs only 600 grams and delivers an incredible 600 Watts - equating to a power density of 25 Watts per cubic inch. The input module can accommodate up to four isolated output modules which can be configured into a high power 5"x 3" single output power supply or a multiple output power supply with up to 8 isolated outputs. Standard features include intelligent fan control providing optimised airflow for various load and temperature conditions, wide output voltage adjust, parallel and series connection of modules and an isolated 5V 1A bias supply. A low noise fan option is available that allows you to use this innovative power supply in even the quietest of environments. The series is approved to latest medical standards and features market leading specifications and design in application support.

MAIN FEATURES

- 600 Watts in 5" x 3" x 1.61"
- User and field configurable
- Wide output voltage adjust range
- Remote current & voltage programming
- Efficiency up to 89%
- Intelligent fan control
- Parallel & series connection of modules
- Standard 5V 1A bias supplyAccurate current sharing
- on of modules IEC/UL6
- IEC/UL60601-1 Ed. 3 & -1-2 Ed. 4 (EMC)
 - 3 Year warranty

• Up to 8 isolated outputs

Low noise option (ML version)

APPLICATIONS

Medical & diagnostic equipment	Telecommunications	Lasers
 Test & Measurement equipment 	 Laboratory & Analysis equipment 	LED lighting
Robotics	• Display	 Retrofit of legacy PSUs
Oil & Gas	Avionics	

CUSTOMER BENEFITS

- Fast time to market
- 24 hrs samples from distribution
- Safety & EMC certified
- World class engineering support
- Proven technologyEliminates custom design costs
- Field replaceable
- Field replaceable
 Low cost of ownership
- Technology consolidationSupplier consolidation
- Redundant manufacturing sites
- Page 1 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com

SPECIFICATIONS

INPUT MODULE SPECIFICATIONS					
Parameter	Details	Min	Typical	Max	Units
AC Input Voltage	Nominal range is 100V _{RMS} to 240V _{RMS}	85		264	V _{RMS}
AC Input Frequency	Contact factory for 400Hz operation.	47	50/60	63	Hz
DC Input Voltage	Not covered by safety approvals. Contact Vox Power.	120		300	V _{DC}
Output Power Rating	De-rate linearly from 600Watts at 120V _{RMS} to 450Watts at 85V _{RMS}			600	Watts
Input Current	600Watts output at 120 V _{RMS} input			6	Amps
Input Current Limit	Maintains power factor		8		Amps
Inrush Current	265V _{RMS} , 25°C (cold start)			20	Amps
Fusing	Live line fused (5x20 Fast acting)			8	Amps
Efficiency	See graphs		86	89	%
No load Power consumption	All outputs fitted and disabled/enabled		21/28		Watts
Power Factor	Typical value for 300 Watts output at 240Vrms input		0.96	0.99	
Holdup	600Watts output at 120V _{RMS} input	17	20	21	mS
UVP	Turn on under voltage protection	78		84	V _{RMS}
Over temperature	Internally monitored.	115		125	°C
Reliability (1)	Input module			1.207	FPMH
	Fan			2.7	FPMH
Warranty	Standard terms and conditions apply			3	Years
Size	133.7 (L) x 77.7 (W) x 41.0 (H). See diagram for tolerance details				mm
Weight	360 + 60 per output module				Grams
Note 1.	30°C base & ambient, 100% load, SR332 Issue 2 Method I, Case 3, Ground, Fixed, Control	led			

GLOBAL SIGNALS SPECIFICATIONS					
Parameter	Details	Min	Typical	Max	Units
Bias Voltage	Two isolated Bias Outputs available	4.8	5	5.2	Volts
Bias Current	Hiccup type current limit	0		1	Amps
AC_OK Voltage	Low output level High output level	0 3.5	0.2 4.5	1 5.2	Volts
AC_OK Current		-10		20	mA
Power Good Voltage	Low output level. internal 10kΩ pull down. High output level. PNP open collector.	0 8	0 10	0 15	Volts
Power Good Current	Open collector output. Current source only. All Slots.			20	mA
Global Inhibit Voltage	Low input level High input level	0 3		1 15	Volts
Global Inhibit Current	5k input impedance.	0.6		3	mA
Inhibit Voltage	Low input level. All slots. High input level. All slots.	0 2.5		1 15	Volts
Inhibit Current	10k input impedance. All slots.	0.25		1.5	mA

OUTPUT MODULE SPECIFICATION SUMMARY

MODEL	Out	put Volta	age	Output	Rated	Peak	Load	Line	Cross	Ripple &	FPMH ⁽¹⁾	Feature
MODEL	Min.	Nom.	Max.	Current	Power	Power	Reg.	Reg.	Reg.	Noise	1170111	Set (2)
OP1	1.5V	5V	7.5V	25A	125W	187.5W	±50mV	±5mV	±10mV	50mV _{PP}	0.5	ABCDEFG
OP2	4.5V	12V	15V	15A	150W	225W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFG
OP3	9V	24V	30V	7.5A	150W	225W	±150mV	±24mV	±48mV	240mV _{PP}	0.5	ABCDEFG
OP4	18V	48V	58V	3.75A	150W	217.5W	±300mV	±48mV	±96mV	480mV _{PP}	0.5	ABCDEFG
OP5	3.3V	12V	15V	5A	2x 75W	2x 75W	±50mV	±12mV	±24mV	240mV _{PP}	0.75	AFG
OP8	23.2V	24V	24.7V	3.125A	2x 75W	2x 75W	±100mV	±24mV	±48mV	480mV _{PP}	0.75	AFG
OPA2 ⁽³⁾	4.5V	12V	15V	25A	300W	375W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFGH
OPA3(3)	9V	24V	30V	15A	300W	450W	±150mV	±24mV	±48mV	$240 mV_{PP}$	0.5	ABCDEFGH
Note 1.	Output r	nodule, 30'	°C base, 10	00% load, SR332	issue 2 Metho	d I. Case 3. Gro	und, Fixed, Co	ontrolled				

Note 1. A = Remote Sense, B = External Voltage control, C = External constant current control, D = Current output signal, E = Current share, F = Over Voltage protection, G = Over temperature protection, H = Dual Slot module

G = Over temperature protection, H = Dual Slot module Note 3. Can only be used with NEVO+600 chassis with date codes from 2048 onwards. eg. 2048C080000 can use A2 or A3 module, 2047C089999 cannot use A2 or A3 module.

	SAFETY SPECIFICATIONS		
Parameter	Details	Max	Units
	Input to Output (2 MOPP). Do not perform test on assembled unit ⁽¹⁾	4000	V _{AC}
	Input to Chassis (1 MOPP)	1500	V _{AC}
Isolation Voltages	Global signals (J2) to Output/Chassis	250	V _{DC}
	Output to Output/Chassis (Standard modules)	250	V _{DC}
Earth Leakage Current	Normal condition, 264Vac, 63Hz, 25°C	300	uA
Touch Leakage Current	Standard modules NC/SFC	20/200	uA
Patient Leakage Current	Standard modules 264Vac, 63Hz, 25°C NC/SFC ⁽²⁾		uA
Note 1. Testing an assembled un	it to 4000V _{AC} may cause damage. Please refer to application note (APN-002) on Vox Power website or contact V	ox Power repres	sentative.

Note 2. Not Applicable

	INSTAL	LATION SPECIFICATIONS	
Parameter	Details	Parameter	Details
Equipment class	I	Flammability Rating	94V-2
Overvoltage category	I	Ingress protection rating	IP10
Material Group	IIIb (indoor use only)	ROHS compliance	2011/65/EU & 2015/863/EU
Pollution degree	2	Intended usage environment	Home Healthcare


		ENVIRONMENTAL SPEC	IFICAT	IONS				
Deverenter	Dataila			Non-Op	erational	Opera	Operational	
Parameter	Details	Details -			Max	Min	Max	Units
Air Temperature	Operational limits subject to appro	priate de-ratings		-40	+85	-20	70	°C
Humidity	Relative, non-condensing			5	95	5	95	%
Altitude				-200	5000	-200	3000	m
Air Pressure				52	106	69	106	kPa
Noise Level	Variable. Measured 1m from fan in	take.		-	-	36	62	dBA
Shock	3000 bumps at 10G (16ms) half sine	e wave						
Vibration	1.5G 10 to 200Hz sine wave, 20G fo	r 15min in 3 axes random vibration						
	EL	ECTROMAGNETIC COMPLIA	NCE –	EMISSIO	NS			
Phenomenon		Basic EMC Standard		Tes	t Details			
Radiated emission	ns, electric field	EN55011/22, FCC		Class	B compliant			
Conducted emissi		EN55011/22, FCC part 15, CISPR 22/11 Class B compliant						
Harmonic Distorti	on	IEC61000-3-2	Compliant					
Flicker & Fluctuati	on	IEC61000-3-3	Compliant					
	EL	ECTROMAGNETIC COMPLIA	ANCE –		TY			
Phenomenon		Basic EMC Standard	Test [Details				
Electrostatic disch	narge	IEC61000-4-2	Test le	vel 4: 15kV ai	r, 8kV contact			
Radiated RF EM fie	elds	IEC61000-4-3	Test Le	evel 3: (10V/m	, 80MHz-2.7G	Hz) sine wave	AM 80% 1kHz	Z
,	om RF wireless communications	IEC61000-4-3	Test levels as per IEC60601-1-2:2014 Table 9					
equipment	· · · · · · · · · · · · · · · · · · ·	15661000 4 4	Tester	· · · · ·	111/1/0)			
Electrical Fast Tran	nsients/dursts	IEC61000-4-4	Test Level 3: (2kV Power, 1kV I/O) 5kHz(ed3) & 100kHz(ed4) Test Level 3: 1kV L-N, 2kV L-E					
Surges		IEC61000-4-5					000/ 1111-	
	bances induced by RF fields	IEC61000-4-6			15 to 80Mhz si	ne wave Alvia	80% IKHZ	
Power Frequency		IEC61000-4-8		vel 4: 30A/m		10- 000/		(a.c. (A))
Voltage Dips		IEC61000-4-11& SEMI-F47-0706 ⁽²⁾	0% 10ms, 0% 20ms, 80% 1s, 80% 10s, 90% continuous (Criterion A) 70% 0.5s, 40% 0.2s (Criterion A at 240V and Criterion B at 100V)					
Voltage interrupti	ions	IEC61000-4-11	0% 250	0/300 cycle a	s per IEC60601	-1-2:2014 (Cr	iterion B)	
Notes: 1. (Criterion A = No degradation of perforr	mance or loss of function.	•					
(Criterion $B = Temporary degradation of the second second$	f performance or loss of function is allo	wed, prov	vided the fun	ction is self-re	coverable.		
Criterion C = Temporary loss of function is allowed but requires operator intervention to recover.								


Tested at nominal range (100V to 240V). Line deratings applied where appropriate.

	AGENCY APPROVALS	
Standard	Details	File
IEC 60601-1:2005 + CORR1 2006 + CORR2: 2007 + A1:2012	Medical electrical equipment Part 1: General requirements for basic safety and essential performance	UL: E316486
EN60601-1:2006 + A11:2011 + A1:2013 + A12:2014	Medical electrical equipment Part 1: General requirements for basic safety and essential performance	
CAN/CSA-C22.2 No. 60601-1 (2008)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance	
ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance	
CE MARK	LVD 2014/35/EU, EMC 2014/30/EU	

CB certificate and report available on request








Temperature

MECHANICAL DIMENSIONS AND MOUNTING SCREWS



CONNECTORS

PINOUTS					
	J1				
Circuit	Details				
1	Live				
2	Earth				
3	Neutral				
	J2				
Circuit	Details				
1	Power good	Slot A			
2	Inhibit	SIOLA			
3	Power good	Slot B			
4	Inhibit	310t D			
5	Power good	Slot C			
6	Inhibit	3101 C			
7	Power good	Slot D			
8	Inhibit	SIDED			
9	Global inhibit				
10	AC OK				
11	+5V 1A bias supply				
12	COM				
	J5 ⁽⁴⁾				

Circuit

4


5


-Sense


+Sense


COM

Voltage control Current control / share / out











REF.	DETAILS	MANUFACTURER	HOUSING	TERMINAL	
J1	MAINS INPUT: 3 Pin, 5.08mm, with Friction Lock, 18-24 AWG	MOLEX	10013036	0008701031	
J2	GLOBAL SIGNALS: 12 Pin, 2mm, without Friction Lock, 24-30 AWG	MOLEX	511101251	0503948051	
J3/4 ⁽¹⁾	OUTPUT POWER TERMINAL: TAB SIZE 6.35mmx0.8mm	VARIOUS		VARIOUS	
J5	OUTPUT SIGNALS: 6 Pin, 1.25mm, with Friction lock, 28-32 AWG	MOLEX	0510210600	0500588000	
Notes					
1. Terminal and wire current rating must exceed maximum short circuit output current. Eg. Output 1 = 25A*1.25 = 31.25Amps					
2. Direct e	equivalents may be used for any connector parts				

:13

-J5

14

3. All cables must be rated 105°C min, equivalent to UL1015

Details

+5V 10mA local bias supply

Pinout is for single output types only 4

DART NI IMRERING SVSTEM

NEVO+ Power Series 💶 NEVO+600 M - 1 1 2 3 - 0 0 0 D Factory Use
Leakage Current Use 0 for unused slots.
Blanking plates will be
M – Medical
Slot A Output No
Slot B Output No Slot C Output No
Our design team will assist with value-add requirement if an application requires standard/non-standard accessories or non-nominal voltage settings.
Once approved, the factory will issue a 3 or 4 digit code for your specific configuration which can be used for all future orders of the same
configuration. When ordering an input unit with no outputs inserted, simply order NEVO+600M

All specifications are believed to be correct at time of publishing. Vox Power Ltd reserves the right to make changes to any of its products and to change or improve any part of the specification, electrical or mechanical design or manufacturing process without notice. Vox Power Ltd does not assume any liability arising out of the use or application of any of its products and of any information to the maximum extent permitted by law. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any products of Vox Power Ltd. VOX POWER LTD DISCLAIMS ALL WARRANTIES AND REPRESENTATIONS OF ANY KIND WHETHER EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, IMPLIED WARRANTIES OF SUITABILITY, FITNESS FOR PURPOSE, MERCHANTABILITY AND NONINFRINGEMENT.

Please consult your local distributor or Vox Power directly to ensure that you have the latest revision before using the product and refer to the latest relevant user manual for further information relating to the use of the product. Vox Power Ltd products are not intended for use in connection with life support systems, human implantations, nuclear facilities or systems, aircraft, spacecraft, military or naval missile, ground support or control equipment used for the purpose of guidance navigation or direction of any aircraft, spacecraft or military or naval missile or any other application where product failure could lead to loss of life or catastrophic property damage. The user will hold Vox Power Ltd harmless from any loss, cost or damage resulting from its breach of these provisions.