
THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

REVISION						
REV.	DESCRIPTION	DATE	APPROVED			



CONTROLLED

NOTES:

- 1. CORDAGE: SJT, 12AWG / 3C , UNSHIELDED, NA COLOR CODE, TEMP. RATING: 60°C, RATING: 125V, 20A, JACKET COLOR: GRAY
- 2. PLUG: NEMA 5-20P HOSPITAL GRADE (CLEAR)
- 3. CONNECTOR: IEC 60320 C19 RIGHT ANGLED (CLEAR)
- 4. APPROVALS: UL, CSA
- 5. ROHSCOMPLIANT
- 6. CAUTION LABEL

Dim range	Tolerance +/-	Qualtek Electronics Corp. PPC DIVISION		
(mm)	(mm)	Part Number:	233106-06	RoH5 Compliant
0-1	0.15	Description: POWER CORD HOSPITAL GRADE		
1-6	0.20			
6-18	0.30		LINUT: :	
18-50	0.50		UNIT: in	
50-120	0.70	Drawn / Date	Checked / Date	Approved / Date
120-250	0.90	RNuti	03-22-18	LY / 03-22-18
250-500	1.20	03-22-18	05-22-10	L1 / 00 22-10