DESCRIPTION APPROVED LTR DATE NOTES: UNLESS OTHERWISE SPECIFIED 1. P/N: DB750BLK HOUSING: BLK GLASS FILLED NYLON. FRAME: NICKEL PLATED BRASS. CONTACT: NICKEL PLATED BeCu OR PHOSPHOR BRONZE. HARDWARE: BLACK OXIDE STEEL WIRE CLAMPING SET SCREW. STAINLESS STEEL HOUSING CLAMPING NUT & SCREW. 2. P/N: Y-1000 HOUSING: BLK MOLDED NYLON HARDWARE: SS SCREW, NUTS & WASHER 3. P/N: XMBLK CENTER O O RED HOUSING: BLK MOLDED NYLON. GROUND O O BLACK CONTACT: GOLD PLATED BeCu. SPRING: STAINLESS STEEL. WIRING SCHEMATIC 4. P/N: XMRED HOUSING: RED MOLDED NYLON. CONTACT: GOLD PLATED BeCu. SPRING: STAINLESS STEEL. **←** 1.73 [43.94] **←** 5. P/N: W-500/0 -/3\GROUND 20AWG TINNED COPPER STRANDING 0.080 [2.03] O.D. PVC INSULATION - 0.852 [21.64] 6. P/N: W-500/2 20AWG TINNED COPPER STRANDING 0.080 [2.03] O.D. PVC INSULATION E-Z-HOOK MADE IN U.S.A. ( 0.10 [2.54] -7. P/N: C-1000/0 9.0 [228.6] 20AWG RG58C/U CABLE ØE-Z-HOOK 0.75 [19.05] · A CENTER 3.0 [76.2] 0.195 [4.95] O.D. PVC INSULATION U.S.A. 1.312 [33.32] **←** 1.6 [40.64] **→** CONTRACT NO. UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES OK DIVISION OF TEKTEST INC. ARCADIA, CA 91006 U.S.A. TOLERANCES ARE: FRACTIONS DECIMALS ANGLES APPROVALS DATE ±N/A  $.XX \pm N/A \pm N/A$ 12/18 DESCRIPTION DRAWN R.L  $A \times \dot{N} \pm \dot{N} / A$ TO XM MICRO DB750 APPROVED BY MATERIAL SEE NOTE RG58 CABLE FINISH SIZE CODE IDENT NO. DRAWING NO. SEE NOTE B |08505| **NEXT ASSY** USED ON SCALE APPLICATION DO NOT SCALE DRAWING SHEET 1 OF 1

THE INFORMATION CONTAINED IN THIS DOCUMENT IS PROPRIETARY TO E-Z-HOOK. THE INFORMATION IS NOT TO BE REPRODUCED WITHOUT PRIOR WRITTEN PERMISSION OF E-Z-HOOK, DIV. OF TEKTEST INC.
MAXIMUM TESTED RATINGS IN A HANDS-FREE CONTROLLED ENVIRONMENT. MAXIMUM RATINGS ARE NOT TO BE USED CONCURRENTLY.

REVISIONS