



All dimensions are in mm; tolerances according to ISO 2768 m-H

Interface

SnapN according to	IEC 61169-16
RPC-N according to	IEC 61169-16

Documents

Application note	AN001 "Calibration Services"
------------------	------------------------------

Material and plating

Connector parts

	Material	Plating
Center contact	CuBe	Gold, min. 1.27 µm, over nickel
Outer contact	Stainless steel	Passivated
Coupling nut	Stainless steel	Passivated
Dielectric	PPE	

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG

Electrical data

Frequency	DC to 11 GHz
Return loss	≥ 36 dB, DC to 4 GHz ≥ 27 dB, 4 GHz to 11 GHz

Mechanical data

	RPC-N 50 Ω	SnapN 50 Ω
Mating cycles	≥ 500	≥ 200
Maximum torque	1.70 Nm	1.70 Nm
Recommended torque	1.10 Nm	1.10 Nm
Engagement force		30 N (typ.)
Disengagement force		30 N (typ.)
Gauge	5.28 mm to 5.32 mm	5.18 mm to 5.26 mm

General standard definition

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Offset Z_0 / Impedance / Z_0	50 Ω
Offset Delay	145.1004 ps
Length (electrical) / Offset Length	43.50 mm
Offset Loss	2.66 GΩ/s
Loss	0.0331 dB/√GHz

Environmental data

Operating temperature range ¹	+20 °C to +26 °C
Rated temperature range of use ²	0 °C to +50 °C
Storage temperature range	-40 °C to +85 °C

RoHS compliant

¹ Temperature range over which these specifications are valid.

² This range is underneath and above the operating temperature range, within the open circuit is fully functional and could be used without damage.

Technical Data Sheet

Rosenberger

Calibration Adaptor
SnapN 50Ω Jack – RPC-N 50Ω Plug

53QK105-S20S3

Declaration of calibration options

Factory Calibration

Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions are reported in an Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Packing

Standard 1 pce in box
Weight 64.6 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Kerstin Herzog	04.06.07	Herbert Babinger	09.11.15	c00	14-1492	Maik Knoll	09.11.15

Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de	Tel. : +49 8684 18-0 Fax : +49 8684 18-1499 Email : info@rosenberger.de	Page 3 / 3
--	--	---------------