



Model 6575 **C E** Alligator clip with Boot for banana plug style test probe.





FEATURES:

- ✓ Complies with International Safety Standard IEC 1010-2-031
- ✓ Alligator clip securely attaches to terminals up to .30" (7,62mm) in diameter.
- ✓ Connects to .166" (4mm), standard banana style test probe. Such as Pomona models 5906A or 6601.

MATERIALS:

Alligator Clip: Nickel plated steel

Boot: Santoprene rubber

RATINGS:

Operating Voltage: 1,000V, Category III

Current: 10 Amperes

Temperature: -13 to 176 F (-25 to 80 C)

ORDERING INFORMATION:

6575-0 (one alligator clip with black boot) 6575-2 (one alligator clip with red boot)

6575-02 (two alligator clips, one red and one black)

All dimensions are in inches. Tolerances (except noted): $.xx = \pm .02$ " (,51 mm), $.xxx = \pm .005$ " (,127 mm). All specifications are to the latest revisions. Specifications are subject to change without notice. Registered trademarks are the property of their respective companies.

Sales: 800-490-2361 Fax: 888-403-3360 Technical Assistance: 800-241-2060 For "Where to Buy" information, visit the Pomona web site at www.pomonaelectronics.com

Rev: 100 D1096780



Model 6575 **C €** Alligator clip with Boot for banana plug style test probe.

^Pomona[®]

The items included with this product are certified to meet IEC 1010-2-031 safety requirements.

Les élements inclus avec ce produit sont certifies pour répondre à des exigences de sûrete du IEC 1010-2-031.

Los items incluidos con este producto se certifican para resolver requisitos de seguridad del IEC 1010-2-031.


 Only IEC 1010 type shrouded test plugs should be used with IEC 1010 type jacks. NEVER INSERT A PROBE TIP WITH BANANA SPRING INTO A TEST ACCESSORY JACK THAT WILL BE USED AS A HAND HELD PROBING DEVICE -Seuls les vérificateurs de tension renforcés de type IEC 1010 devraient être utilisés avec des jacks de type IEC 1010. NE JAMAIS INSERER DE BOUT DE SONDE AVEC RESSORT BANANE DANS UN JACK D'ESSAI D'ACCESSOIRES QUI SERA UTILISE COMME DISPOSITIF DE SONDAGE A MAIN.

-Se deben usar sólo tapones de prueba reforzados del tipo IEC 1010 con receptáculos de tipo IEC 1010. NUNCA INTRODUZCA UNA CLAVIJA CON UN MUELLE TIPO BANANA EN UN RECEPTACULO DE PRUEBA QUE A SU VEZ SE USE COMO APARATO DE PRUEBA MANUAL.





Caution! Refer to Accompanying Documents. Attention! Se Reporter aux Documentsjoints.; Peligro! Consulte los Documentos Adjuntos.





Caution! Risk of Electric Shock. Attention! Risque de Choc Électrique. ¡Peligro! Riesgo de Sacudida Electrica.

Overvoltage Category I, II, III or IV.
CAT Surtension Catégorie I, II, III ou IV.
Sobretensión de Categoria I, II III o IV.

Pomona Electronics (E

All dimensions are in inches. Tolerances (except noted): $.xx = \pm .02$ " (,51 mm), $.xxx = \pm .005$ " (,127 mm). All specifications are to the latest revisions. Specifications are subject to change without notice. Registered trademarks are the property of their respective companies.

Sales: 800-490-2361 Fax: 888-403-3360 Technical Assistance: 800-241-2060 For "Where to Buy" information, visit the Pomona web site at www.pomonaelectronics.com

Rev: 100 D1096780