

Manual Supplement

Manual Title:	123B/124B/125B Users	Supplement Issue:	4
Print Date:	January 2016	Issue Date:	9/17
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 257


On page 54, replace step 9 with:

9. Push **ENTER** and use **▼** and **ENTER** to set the percentage of deviation for time (relative to 1 period) and the amplitude deviation (relative to peak-to-peak) for waveforms. Threshold can be used for signals up to 4 kHz.

Change #2, 380

On the Cover, add the **BC** Symbol.

On page 4, add the following to the **Symbols** table:

	Conforms to the Appliance Efficiency Regulation (California Code of Regulations, Title 20, Sections 1601 through 1608), for small battery charging systems.
-----------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------