





Directions:

Remove booties from ACL 306 Two-Point Probe. Place pogo pins on desired decade. Depress tester while pressing test button. The LED for corresponding decade will illuminate.



Product# 307: includes board checker and calibration certificate

Product# 306: includes tester, extra set of pins, one set of rubber boots, and calibration certificate

ACL 307 Checker Board

Calibration Checker for the ACL 306

The ACL 307 Checker Board offers quick and easy verification of the ACL 306 handheld Two-Point Resistance Tester.

The Checker Board is equipped with logarithmic landing spots for each decade on the 306 tester's scale. Each pad is designed with the correct spacing to accommodate the tester's two pogo-pins. Every board is serial numbered and calibrated.

Ideal for the following applications:

Verifying ACL 306 Two-Point Resistance Tester

FEATURES:

- Measuring range: 10e3 10e10 ohms
- Measurements over 10e10 ohms will indicate ≥ 10e11
- Metal pads with 1% resistor
- Dimensions: 5" x 3.25" x 0.19"
- Made in USA



ACL**Staticide**

840 W. 49th Place · Chicago, IL 60609 800.782.8420 · 847.981.9212 · fax 847.981.9278 info@aclstaticide.com · www.aclstaticide.com



ACL, Inc. has been certified by DQS Inc. to ISO-9001 : 2015 Certificate NO. 10002746